



*Topic Expert Group: Medical care and clinical practice*

**Neurological monitoring in the high-risk infant: Near-infrared spectroscopy (NIRS)**

Pellicer A, Hellström-Westas L, Zimmermann L, Buonocore G, Dudink J, Gressens P

*Target group*

- Term and preterm infants at risk for brain injury:
  - Infants with hypoxic-ischaemic encephalopathy (HIE)
  - Infants with encephalopathy for other causes (e.g. metabolic)
  - Infants with suspected or verified seizures
  - Infants requiring intensive care and/or surgery
  - Infants with suspected/confirmed congenital central nervous system (CNS) anomalies
- Parents

*User group*

Healthcare professionals, neonatal units, hospitals, and health services

*Statement of standard*

In order to improve evaluation and outcomes of newborn infants at risk of brain injury, management includes neurological monitoring using a structured, age-appropriate neurological assessment and a range of devices to evaluate brain haemodynamics, oxygen transport, brain function, and imaging, as required.

*Rationale*

Infants requiring neonatal intensive care constitute a high-risk population for developing brain injury, particularly full term and preterm infants exposed to hypoxia-ischaemia, CNS infections, or with congenital anomalies.

In the first hours after birth, there is imbalance between blood flow and oxygen supply to the brain due to haemodynamic adaptation during transitional circulation, particularly in the very preterm infant. (1) Low and fluctuating cerebral blood flow are associated with adverse outcomes. (2,3) Experimental models and observational studies confirm that both hyper- and hypoxaemia may cause irreversible brain injury. (4–6) The vulnerability of this population, the severity of underlying clinical conditions, and the complexity of care make continuous, cot-side, and non-invasive monitoring tools valuable. Near-infrared spectroscopy (NIRS) derived regional tissue oxygen saturation of haemoglobin (rStO<sub>2</sub>) is an absolute value, which corresponds to mixed blood saturation, used in the clinical setting as a surrogate measure for venous oxygen saturation (SvO<sub>2</sub>). (7) Indirect assessment of cerebral blood flow has been shown to correlate with rStO<sub>2</sub>. (8) This non-invasive, continuous monitoring system may help to adjust interventions that have effects on blood and oxygen supply to the brain. (9) Bilateral brain monitoring may detect differential perfusion between hemispheres.



## Benefits

### Short-term benefits

- Reduced burden of cerebral hypo- and hyperoxia in preterm infants in the first 72 h after birth (10,11)
- Improved neuroprotection after asphyxia using combined NIRS and MRI measurements of brain perfusion (12)
- Improved maintenance of theoretically safe cerebral oxygenation levels in infants with congenital heart defects (13)

### Long-term benefits

- Reduced all-cause mortality in extremely preterm infants (10)
- Improved long-term outcomes in extremely preterm infants (14)

### Components of the standard

Component	Grading of evidence	Indicator of meeting the standard
<b>For parents and family</b>		
1. Parents are informed by healthcare professionals about the role of near-infrared spectroscopy (NIRS) monitoring.	B (High quality)	Patient information sheet
<b>For healthcare professionals</b>		
2. A unit guideline on neurological monitoring including NIRS is adhered to by all healthcare professionals, to include <ul style="list-style-type: none"> <li>• Newborn infants during resuscitation at birth (<math>\leq 15</math> min) (11,15,16)</li> <li>• Extremely preterm infants in the first 72 h after birth (9,10,17)</li> <li>• Asphyxiated newborn infants undergoing therapeutic hypothermia (18,19)</li> <li>• Infants undergoing surgery with cardio-pulmonary bypass (13,20–22)</li> </ul>	A (High quality) B (High quality)	Guideline
3. Training on NIRS monitoring is attended by all responsible healthcare professionals. (9,17,20,21,23)	A (High quality) B (High quality)	Training documentation
4. Teams with a focus of interest on neuro-critical care, including neonatologists, neurologists, neuro-physiologists, nurses, radiologists, radiographers, and physicists are	B (High quality)	Guideline



established.

For neonatal unit		
5. A unit guideline on neurological monitoring including NIRS is available and regularly updated including standardised operational procedures. (7,9,17,23)	A (High quality) B (High quality)	Guideline
For hospital		
6. Training on NIRS monitoring is ensured. (7,17,20,21,23)	A (High quality) B (High quality)	Training documentation
7. Facilities for NIRS monitoring are provided.	B (High quality)	Audit report
8. An interdisciplinary team for neuro-critical care of high-risk infants in the NICU is supported.	B (Moderate quality)	Audit report
For health service		
9. High-risk infants are transferred to NICUs with appropriate neuro-monitoring systems and expertise. (24–26)	A (High quality)	Audit report, guideline

### *Where to go – further development of care*

Further development	Grading of evidence
For parents and family N/A	
For healthcare professionals	
<ul style="list-style-type: none"> <li>Monitor perioperative NIRS in infants with non-cardiac complex neonatal surgery. (27,28)</li> </ul>	A (Low quality)
For neonatal unit N/A	
For hospital N/A	
For health service N/A	

### *Getting started*

Initial steps
For parents and family
<ul style="list-style-type: none"> <li>Parents are verbally informed by healthcare professionals about the role of NIRS.</li> </ul>



#### For healthcare professionals

- Attend training on NIRS monitoring.
- Identify leading healthcare professionals with a focus of interest on neonatal neurological monitoring.

#### For neonatal unit

- Develop and implement a unit guideline on neurological monitoring including NIRS.
- Develop parental information material about NIRS monitoring also including parent perspectives.
- Provide resources for specific training on NIRS monitoring.

#### For hospital

- Support healthcare professionals to participate in training on NIRS monitoring.

#### For health service

- Create systems to effectively transfer high-risk infants to NICUs with appropriate neuro-monitoring systems and expertise.

### *Description*

The NIRS sensor is placed at the forehead avoiding cavities, superior sagittal sinus, intra or extra-cranial huge blood collections, or vascular malformations, if known. Scalp oedema will also influence the quality of the NIRS signal. In the smallest newborn infants and those with poor perfusion states sensor position is rotated to avoid tissue injury related to compression or heat. (7,17)

Commercial NIRS devices incorporate similar technology but different wavelengths and computational algorithms translating changes in light absorption into  $rStO_2$  absolute values. (7) Systematic approach has evidenced huge differences in  $rStO_2$  according to device or probe (23,29), so that device-specific reference ranges or limits have to be used.

*Neonatal resuscitation after birth:* Clinical assessment of the newborn infant carries high inter-observer variability particularly when scoring preterm or term infants in need of resuscitation. (30) Oxygen saturation targeting and the use of supplemental oxygen during transition remain controversial topics. (31) The use of pulse oximetry or heart rate monitoring during resuscitation has not led to improvements on the short or long-term outcomes. (32)  $rStO_2$  and fractional oxygen extraction reference ranges and percentile charts for the interpretation of cerebral oxygenation during immediate transition to avoid hypo- and hyperoxia of the brain during resuscitation appears promising. (11,15,16) Yet, routine interventions based on  $rStO_2$  during resuscitation need development and evaluation.

*Extremely low gestational age newborn infants:* Recent studies have shown an association of cerebral  $rStO_2$  levels and clinical outcomes. (33) Low  $rStO_2$  on the first day of life associates surrogated measures of compromised systemic blood flow and risk of intraventricular haemorrhage. (34) Impaired cerebral blood flow autoregulation assessed by NIRS and arterial blood pressure monitoring associates abnormal systemic (and cerebral) blood flow distribution, death and severe brain injury. (35,36) Cerebral oxygenation can be stabilised in the preterm infant during the first 72 hours from birth by the combined use of  $rStO_2$ -NIRS monitoring and a pathophysiological, brain oriented treatment guideline with no record of severe adverse events. (9,10) The quality of evidence supporting some of the listed statements in the intervention



algorithm is generally low, however, are all routinely used in clinical care of these patients. (9) Although important early surrogate outcomes, such as aEEG at day 3 of postnatal life or neuroimaging, did not significantly differ between the study groups (37,38), post hoc analyses showed that early burden of cerebral hypoxia was significantly associated with low brain electrical activity and severity of intracranial haemorrhage. (14) So far, definitive evidence of benefit for improvement of long-term clinical outcomes is needed as the technology is not cheap, requires manipulation and additional staff time, and may have unwanted effects. (10)

*HIE:* Cerebral hypoperfusion during the first hours after birth is followed by hyperperfusion, even during treatment with moderate hypothermia. Potential differences according to the severity of brain injury (moderate vs severe) have been identified. (12,18) NIRS measurements of oxygenation and MRI measurements of brain perfusion show good correlation. (12) However, the predictive capacity of NIRS changes lacks consistency. (18,19) As yet, widespread recommendation of NIRS monitoring to guide important clinical decisions in asphyxiated newborn infants cannot be made.

*Congenital heart disease (CHD):* NIRS may be a useful adjunct particularly during cardiopulmonary bypass to optimise perfusion. NIRS-derived measures of systemic oxygen balance correlate with global circulatory measures and biochemical indicators of shock. (20) Algorithms have been developed to guide interventions based on rStO<sub>2</sub> values during the perioperative period. (21) However, the current literature on the use of NIRS alone does not demonstrate improvement in neurologic outcome. (22) Prospective data evaluating NIRS findings and relevant outcomes in this population difficult to compare, because of the variable disease physiology, variable baseline values, and small sample sizes. These issues prevent extrapolation to wider CHD population.

Other complex surgical procedures conducted during the neonatal period, such as congenital diaphragmatic hernia or esophageal atresia (27,28), might be additional scenarios where NIRS may play a role to guide surgeons and anesthetists during the intervention procedures.

## Source

1. Kluckow M, Seri I. Clinical presentations of neonatal shock: The very low birth weight neonate during the first postnatal day. *Hemodynamics Cardiol.* 2012 Jan 1;237–67.
2. Meek JH, Tyszczuk L, Elwell CE, Wyatt JS. Low cerebral blood flow is a risk factor for severe intraventricular haemorrhage. *Arch Dis Child - Fetal Neonatal Ed.* 1999 Jul 1;81(1):F15–8.
3. Hunt RW, Evans N, Rieger I, Kluckow M. Low superior vena cava flow and neurodevelopment at 3 years in very preterm infants. *J Pediatr.* 2004 Nov 1;145(5):588–92.
4. Kurth CD, McCann JC, Wu J, Miles L, Loepke AW. Cerebral oxygen saturation-time threshold for hypoxic-ischemic injury in piglets. *Anesth Analg.* 2009 Apr;108(4):1268–77.
5. Collins MP, Lorenz JM, Jetton JR, Paneth N. Hypocapnia and other ventilation-related risk factors for cerebral palsy in low birth weight infants. *Pediatr Res.* 2001 Dec;50(6):712–9.
6. Klinger G, Beyene J, Shah P, Perlman M. Do hyperoxaemia and hypocapnia add to the risk of brain injury after intrapartum asphyxia? *Arch Dis Child Fetal Neonatal Ed.* 2005 Jan;90(1):F49-52.



7. Pellicer A, Bravo M del C. Near-infrared spectroscopy: a methodology-focused review. *Semin Fetal Neonatal Med.* 2011 Feb;16(1):42–9.
8. Moran M, Miletin J, Pichova K, Dempsey EM. Cerebral tissue oxygenation index and superior vena cava blood flow in the very low birth weight infant. *Acta Paediatr Oslo Nor* 1992. 2009 Jan;98(1):43–6.
9. Pellicer A, Greisen G, Benders M, Claris O, Dempsey E, Fumagalli M, et al. The SafeBoosC phase II randomised clinical trial: a treatment guideline for targeted near-infrared-derived cerebral tissue oxygenation versus standard treatment in extremely preterm infants. *Neonatology.* 2013;104(3):171–8.
10. Hyttel-Sorensen S, Pellicer A, Alderliesten T, Austin T, van Bel F, Benders M, et al. Cerebral near infrared spectroscopy oximetry in extremely preterm infants: phase II randomised clinical trial. *BMJ.* 2015 Jan 5;350(jan05 2):g7635–g7635.
11. Pichler G, Urlesberger B, Baik N, Schwabegger B, Binder-Heschl C, Avian A, et al. Cerebral Oxygen Saturation to Guide Oxygen Delivery in Preterm Neonates for the Immediate Transition after Birth: A 2-Center Randomized Controlled Pilot Feasibility Trial. *J Pediatr.* 2016 Mar;170:73–78.e4.
12. Wintermark P, Hansen A, Warfield S, Dukhovny D, Soul J. Near-Infrared Spectroscopy versus Magnetic Resonance Imaging To Study Brain Perfusion in Newborns with Hypoxic-Ischemic Encephalopathy Treated with Hypothermia. *NeuroImage.* 2014 Jan 15;85(0 1):287–93.
13. Hoffman GM, Brosig CL, Mussatto KA, Tweddell JS, Ghanayem NS. Perioperative cerebral oxygen saturation in neonates with hypoplastic left heart syndrome and childhood neurodevelopmental outcome. *J Thorac Cardiovasc Surg.* 2013 Nov;146(5):1153–64.
14. Plomgaard AM, Alderliesten T, Austin T, Bel F van, Benders M, Claris O, et al. Early biomarkers of brain injury and cerebral hypo- and hyperoxia in the SafeBoosC II trial. *PLOS ONE.* 2017 Mar 22;12(3):e0173440.
15. Pichler G, Binder C, Avian A, Beckenbach E, Schmölzer GM, Urlesberger B. Reference ranges for regional cerebral tissue oxygen saturation and fractional oxygen extraction in neonates during immediate transition after birth. *J Pediatr.* 2013 Dec;163(6):1558–63.
16. Baik N, Urlesberger B, Schwabegger B, Schmölzer GM, Avian A, Pichler G. Cerebral haemorrhage in preterm neonates: does cerebral regional oxygen saturation during the immediate transition matter? *Arch Dis Child-Fetal Neonatal Ed.* 2015;100(5):F422–F427.
17. Riera J, Hyttel-Sorensen S, Bravo MC, Cabañas F, López-Ortego P, Sanchez L, et al. The SafeBoosC phase II clinical trial: an analysis of the interventions related with the oximeter readings. *Arch Dis Child Fetal Neonatal Ed.* 2016 Jul;101(4):F333–338.
18. Toet MC, Lemmers PMA, Schelven LJ van, Bel F van. Cerebral Oxygenation and Electrical Activity After Birth Asphyxia: Their Relation to Outcome. *Pediatrics.* 2006 Feb 1;117(2):333–9.
19. Ancora G, Maranella E, Grandi S, Sbravati F, Coccolini E, Savini S, et al. Early predictors of short term neurodevelopmental outcome in asphyxiated cooled infants. A combined brain amplitude integrated electroencephalography and near infrared spectroscopy study. *Brain Dev.* 2013 Jan;35(1):26–31.
20. Tweddell JS, Ghanayem NS, Hoffman GM. Pro: NIRS is “standard of care” for postoperative management. *Semin Thorac Cardiovasc Surg Pediatr Card Surg Annu.* 2010;13(1):44–50.
21. Denault A, Deschamps A, Murkin JM. A proposed algorithm for the intraoperative use of cerebral near-infrared spectroscopy. *Semin Cardiothorac Vasc Anesth.* 2007 Dec;11(4):274–81.



22. Hirsch JC, Charpie JR, Ohye RG, Gurney JG. Near-infrared spectroscopy: what we know and what we need to know--a systematic review of the congenital heart disease literature. *J Thorac Cardiovasc Surg.* 2009 Jan;137(1):154–9, 159-12.
23. Hyttel-Sorensen S, Sorensen LC, Riera J, Greisen G. Tissue oximetry: a comparison of mean values of regional tissue saturation, reproducibility and dynamic range of four NIRS-instruments on the human forearm. *Biomed Opt Express.* 2011 Oct 6;2(11):3047–57.
24. Phibbs CS, Baker LC, Caughey AB, Danielsen B, Schmitt SK, Phibbs RH. Level and volume of neonatal intensive care and mortality in very-low-birth-weight infants. *N Engl J Med.* 2007;356(21):2165–2175.
25. Tucker J, UK Neonatal Staffing Study Group. Patient volume, staffing, and workload in relation to risk-adjusted outcomes in a random stratified sample of UK neonatal intensive care units: a prospective evaluation. *Lancet Lond Engl.* 2002 Jan 12;359(9301):99–107.
26. Hannan EL, Racz M, Kavey RE, Quaegebeur JM, Williams R. Pediatric cardiac surgery: the effect of hospital and surgeon volume on in-hospital mortality. *Pediatrics.* 1998 Jun;101(6):963–9.
27. Bishay M, Giacomello L, Retrosi G, Thyoka M, Nah SA, McHoney M, et al. Decreased cerebral oxygen saturation during thoracoscopic repair of congenital diaphragmatic hernia and esophageal atresia in infants. *J Pediatr Surg.* 2011 Jan;46(1):47–51.
28. Bishay M, Giacomello L, Retrosi G, Thyoka M, Garriboli M, Brierley J, et al. Hypercapnia and acidosis during open and thoracoscopic repair of congenital diaphragmatic hernia and esophageal atresia: results of a pilot randomized controlled trial. *Ann Surg.* 2013 Dec;258(6):895–900.
29. Sorensen LC, Greisen G. Precision of measurement of cerebral tissue oxygenation index using near-infrared spectroscopy in preterm neonates. *J Biomed Opt.* 2006 Oct;11(5):54005.
30. Bashambu MT, Whitehead H, Hibbs AM, Martin RJ, Bhola M. Evaluation of interobserver agreement of apgar scoring in preterm infants. *Pediatrics.* 2012 Oct;130(4):e982-987.
31. Dawson JA, Kamlin COF, Vento M, Wong C, Cole TJ, Donath SM, et al. Defining the reference range for oxygen saturation for infants after birth. *Pediatrics.* 2010 Jun;125(6):e1340-1347.
32. Dawson JA, Morley CJ. Monitoring oxygen saturation and heart rate in the early neonatal period. *Semin Fetal Neonatal Med.* 2010 Aug;15(4):203–7.
33. Alderliesten T, Lemmers PMA, van Haastert IC, de Vries LS, Bonestroo HJC, Baerts W, et al. Hypotension in preterm neonates: low blood pressure alone does not affect neurodevelopmental outcome. *J Pediatr.* 2014 May;164(5):986–91.
34. Noori S, McCoy M, Anderson MP, Ramji F, Seri I. Changes in cardiac function and cerebral blood flow in relation to peri/intraventricular hemorrhage in extremely preterm infants. *J Pediatr.* 2014 Feb;164(2):264-270-3.
35. Riera J, Cabañas F, Serrano JJ, Bravo MC, López-Ortego P, Sánchez L, et al. New time-frequency method for cerebral autoregulation in newborns: predictive capacity for clinical outcomes. *J Pediatr.* 2014 Nov;165(5):897–902.e1.
36. Riera J, Cabañas F, Serrano JJ, Madero R, Pellicer A. New developments in cerebral blood flow autoregulation analysis in preterm infants: a mechanistic approach. *Pediatr Res.* 2016 Mar;79(3):460–5.
37. Plomgaard AM, van Oeveren W, Petersen TH, Alderliesten T, Austin T, van Bel F, et al. The SafeBoosC II randomized trial: treatment guided by near-infrared spectroscopy reduces cerebral hypoxia without changing early biomarkers of brain injury. *Pediatr Res.* 2016 Apr;79(4):528–35.



european standards of  
care for newborn health

38. Plomgaard AM, Hagmann C, Alderliesten T, Austin T, van Bel F, Claris O, et al. Brain injury in the international multicenter randomized SafeBoosC phase II feasibility trial: cranial ultrasound and magnetic resonance imaging assessments. *Pediatr Res.* 2016 Mar;79(3):466–72.

First edition, November 2018

*Lifecycle*

5 years/next revision: 2023

*Recommended citation*

EFCNI, Pellicer A, Hellström-Westas L et al., European Standards of Care for Newborn Health: Neurological monitoring in the high-risk infant: Near-infrared spectroscopy (NIRS). 2018.